



Calibration Form

Chillfoodie Manufacturing Inc. / SP1001051 / 20 Aug 2025

Complete

Score	0 / 1 (0%)	Flagged items	1	Actions	1
Site	Chillfoodie Manufacturing Inc.				
Equipment Tag	SP1001051				
Date and Time of Calibration	20.08.2025 08:27 PST				
Technical Person (Full Name)	John Armanio				
Location	California, USA (36.778261, -119.4179324)				

Flagged items & Actions

1 flagged, 1 action

Flagged items

1 flagged, 1 action

Inspection / Equipment

Is the equipment calibrated? (Please attach photo evidence)

No

The temperature is higher than the set standard.



Photo 2

To do | Priority: Medium | Due: 27.08.2025 08:31 PST | Created by: SafetyCulture Staff

Calibrate this equipment

Other actions

0 actions

Inspection

1 flagged, 1 action, 0 / 1 (0%)

Reference

Reference design

Portable Stem Thermometer

Add a photo of the reference design



Photo 1

Tolerance Limits

Range: -40 to 204°C (-40 to 400°F),
Stem Length: 12 Inch

Manufacturer

TempStem

Equipment

1 flagged, 1 action, 0 / 1 (0%)

Model Number:

XZ15126YA

Is the equipment calibrated? (Please attach photo evidence)

No

The temperature is higher than the set standard.



Photo 2

To do | Priority: Medium | Due: 27.08.2025 08:31 PST | Created by: SafetyCulture Staff

Calibrate this equipment

Completion

Overall Recommendations

Calibrate Portable Stem
Thermometer located at Bay 2 as
soon as possible.

Next Calibration Date (Please set this to schedule tab)

03.09.2025

Technical Person (Name and Signature)

John Armanio
20.08.2025 08:32 PST

Media summary



Photo 1



Photo 2